

Two types of loading ports are available, one for standard needles and one for needleless Luer fittings.

The air vent has a microbial filter.

The foil cover on the septum of the retrieval port can be removed easily for sample retrieval.

Sealing the tubing creates a closed system.



CellSeal[®]

©2018 COOK